

EC Declaration of Conformity

Document no. DZ. SG.25.Smart.Z2.ENG

Manufacturer: **APLISENS S.A.**,
ul. Morelowa 7, 03-192 Warszawa

We hereby acknowledge, that:

Smart hydrostatic level probes models **SGE-25.Smart, SGE-25S.Smart, SGE-25C.Smart** in all versions comply with directive

- **EMC – 2004/108/WE**

The following standard was applied PN-EN 61326-1:2006

Smart hydrostatic level probes models **SGE-25.Smart, SGE-25S.Smart, SGE-25C.Smart** in Intrinsically Safe version comply with directive.

- **ATEX – 94/9/WE**

Level probes in Intrinsically Safe version are marked with the following certificate marking:

II 1G
Ga Ex ia IIC T4/T5/T6
I M1, Ex ia I

Certificate no. KDB 09ATEX008

Certificate was issued by Notified Body no. 1453 Central Mining Institute,
Plac Gwarków 1, 40-166 Katowice.

The following standards were applied: PN-EN60079-0:2006, PN-EN60079-11:2007,
PN-EN60079-26:2007, PN-EN50303:2004.

- **Marine version**

Smart hydrostatic level probes models **SGE-25.Smart, SGE-25S.Smart, SGE-25C.Smart** in marine version comply with regulations of Det Norske Veritas to be used for ships, vessels and oil rigs and are marked with the following certificate marking:

DNV No. A-11311

and class of installation standards: temperature: class **C**, humidity: class **B**,
vibrations: class **B**, EMC class: **B**, housing: class **D**.

Certificate was issued by Det Norske Veritas, Veritasveien 1, 1322 Hovik, Norway

Warsaw, 15th October 2009

Piotr Pawelec

P. Pawelec
Quality Manager