



DET NORSKE VERITAS

TYPE APPROVAL CERTIFICATE

CERTIFICATE NO. A-11311
This Certificate consists of 3 pages

This is to certify that the
Hydrostatic level probe

with type designations
**SGE-25Smart (SG-25Smart); SGE-25S.Smart
(SG-25S.Smart); SGE-25 (SG-25); SGE-25S (SG-25S)
SGE-25C (SG-25C)**


Manufactured by
"APLISENS" S.A.
Warszawa, Poland

is found to comply with
Det Norske Veritas' Rules for Classification of Ships, High Speed & Light Craft and
Det Norske Veritas' Offshore Standards

Application
Location classes:

Temperature	Humidity	Vibration	EMC	Enclosure
C	B	B	B	D



Place and date
Høvik, 2009-05-29
for DET NORSKE VERITAS AS


Odd Magne Nesvåg
Head of Section



Local Office
DNV Sopot

This Certificate is valid until
2013-06-30


Andrzej Gdaniec
Surveyor 

Notice: This Certificate is subject to terms and conditions overleaf. Any significant change in design or construction may render this Certificate invalid.
The validity date relates to the Type Approval Certificate and not to the approval of equipment/systems installed.

If any person suffers loss or damage which is proved to have been caused by any negligent act or omission of Det Norske Veritas, then Det Norske Veritas shall pay compensation to such person for his proved direct loss or damage. However, the compensation shall not exceed an amount equal to ten times the fee charged for the service in question, provided that the maximum compensation shall never exceed USD 2 million. In this provision "Det Norske Veritas" shall mean the Foundation Det Norske Veritas as well as all its subsidiaries, directors, officers, employees, agents and any other acting on behalf of Det Norske Veritas.



Cert. No.: A-11311
File No.: 893.30
Job Id.: 262.1-006045-1

Product description

The following type designations are included in the certificate:

Hydrostatic level probe SGE-25Smart (SG-25Smart)
Hydrostatic level probe SGE-25S.Smart (SG-25S.Smart)
Hydrostatic level probe SGE-25 (SG-25)
Hydrostatic level probe SGE-25S (SG-25S)
Hydrostatic level probe SGE-25C (SG-25C)

Application/Limitation

The Type Approval covers hardware listed under Product description. When the hardware is used in applications to be classed by DNV, documentation for the actual application is to be submitted for approval by the manufacturer of the application system in each case. Reference is made to DNV Rules for Ships Pt.4 Ch.9 Control and Monitoring Systems.

Ex-certification is not covered by this certificate. Application in hazardous area to be approved in each case according to the Rules and Ex-Certification/ Special Condition for Safe Use listed in valid Ex-certificate issued by a notified/recognized Certification Body.

The cable installation has to be carried out according to the rules for classification of ships with respect to routing, fixing, protection, penetration of decks / bulkheads and the use of cable pipes. Fixing of cables above sensor shall be provided.

Type Approval documentation

User's Manual	DTR.SG...04
APLISENS Technical specification	Dated 27.08.08
Drawings	SG25-A012-TA; SG25-A004-TA SG25-S005-TA; SG25-S002-TA
APLISENS Product Catalogue	Edition 2008/09
IEL (The Electrotechnical Institute) Gdansk Test Reports	No. 052/LMC-966/2009, dated 25.03.2009 No. 025/LBS-966/2009, dated 03.04.2009 No. 054/LMC-966/2009, dated 25.03.2009 No. 022/LBS-966/2009, dated 03.04.2009 No. 055/LMC-966/2009, dated 25.03.2009 No. 018/LBS-966/2009, dated 31.03.2009 No. 053/LMC-966/2009, dated 25.03.2009 No. 023/LBS-966/2009, dated 03.04.2009 No. 051/LMC-966/2009, dated 25.03.2009 No. 021/LBS-966/2009, dated 03.04.2009



Cert. No.: A-11311
File No.: 893.30
Job Id.: 262.1-006045-1

Tests carried out

Applicable tests according to DNV Standard for Certification No. 2.4, April 2006.

Certificate retention survey

The scope of the retention/renewal survey is to verify that the conditions stipulated for the type are complied with, and that no alterations are made to the product design or choice of systems, software versions, components and/or materials.

The main elements of the survey are:

- Ensure that type approved documentation is available
- Inspection of factory samples, selected at random from the production line (where practicable)
- Review of production and inspection routines, including test records from product sample tests and control routines
- Ensuring that systems, software versions, components and/or materials used comply with type approved documents and/or referenced system, software, component and material specifications
- Review of possible changes in design of systems, software versions, components, materials and/or performance, and make sure that such changes do not affect the type approval given
- Ensuring traceability between manufacturer's product type marking and the type approval certificate

Retention survey is to be performed at least every second year and at renewal of this certificate.

END OF CERTIFICATE